Supplier's Declaration of Conformity

Place of Issue: Nemko USA, Inc.

Date of Issue: 1/29/08

Boston Scientific located at ***4100 Hamline Ave N., St. Paul, MN 55112*** in the United States of America hereby certifies that the ***BSC Model 6476*** bearing labeling identification number ***US:ESCMM00BBSC6476*** complies with the Federal Communications Commission's ("FCC") Rules and Regulations 47 CFR Part 68, and the Administrative Council on Terminal Attachments ("ACTA")-adopted technical criteria **TIA-968-A, Telecommunications – Telephone Terminal Equipment -Technical Requirements for Connection of Terminal Equipment To the Telephone Network, January 2003, TIA-968-A-1 July 2003, TIA-968-A-2 January 2004 & TIA-968-A-3, February 2005, TIA-968-A-4 Nov 2006, TIA-1096 Aug 2006.**

Bob Gaskill

Director, Latitude Embedded Systems
Advanced Patient Management, CRM
Boston Scientific Corporation
4100 Hamline Ave., N
St. Paul, MN 55112-5798